





# CERTIFICATE FOR CHINA COMPULSORY PRODUCT CERTIFICATION



**CERTIFICATE NO: 2021322315003731**

**APPLICANT:** Minco Products, Inc.  
**ADDRESS:** 7300 Commerce Lane NE, Minneapolis, MN 55432, United States of America

**MANUFACTURER:** Minco Products, Inc.  
**ADDRESS:** 7300 Commerce Lane NE, Minneapolis, MN 55432, United States of America

**FACTORY:** Minco Products, Inc.  
**ADDRESS:** 7300 Commerce Lane NE, Minneapolis, MN 55432, United States of America

**PRODUCTNAME:** Resistance Temperature Detector

**SERIES,SPECIFICATION,MODEL:** B215602

**STANDARDS:** GB 3836.1-2010、GB 3836.3-2010、GB 3836.4-2010、GB 3836.8-2014

This is to certify that the above mentioned product(s)complies with the requirements of implementation rules for compulsory certification (REFNO. CNCA-C23-01:2019).

**Valid from:**February 7, 2021

**Valid until:** February 6, 2026

**Date of original certification:**February 7, 2021

**The validity of this certificate is subject to positive result of the regular follow up inspection by issuing certification body until the expiry date.**

This certificate is available through CNCA's website: [www.cnca.gov.cn](http://www.cnca.gov.cn)



APPROVAL:

Guo AiHua



**Shanghai Inspection and Testing Institute of Instruments and Automation Systems Co., Ltd.**

<http://www.sitiias.com.cn>

Building 9,103 Cao Bao Road, Shanghai 200233,China

Tel: +86 21 64510844

s 0004511



# 中国国家强制性产品认证证书



证书编号：2021322315003731

## 附件

产品名称：电阻温度检测器

型号规格：B215602-*abcd*

*a*代表有无接头，可为A或B

*b*代表结构类型，可为1或2或3

*c*代表热缩管和探针帽选项，可为A、B、C或D

*d*代表热缩管和探针帽材质，可为0、1或2

防爆标志：Ex ia IIC T1~T6/T570°C/ T582°C/ T610°C Ga

Ex e IIC T1~T6/T570°C/ T582°C/ T610°C Gb

Ex nA IIC T1~T6/T570°C/ T582°C/ T610°C Gc

电气参数：当产品为本安型时，本安参数为 $U_i=30V$ ， $P_i=0.4W$ ， $C_i=28pF/m$ ， $L_i=1.33uH/m$ 。  
当产品为增安型和nA型时，电气参数为 $U \leq 30V$ ， $P \leq 0.4W$ 。

批准：



## 上海仪器仪表自控系统检验测试所有限公司